

VERIFICATION OF CALIBRATION REPORT

MAR 23 2010

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Name <u>JAMES HOELSCHER</u>	ID# <u>5343</u> Date <u>3-11-10</u>
A Agency <u>HOOPER BAY POLICE DEPT.</u>	Phone # <u>758-4615</u>
Instrument Location <u>Hooper Bay</u>	
Alco S/N <u>X301526</u> Target Value <u>.080</u> High Pressure <u>450 PSI</u>	
B Alco Test Value Average <u>.079</u>	1 st Alco <u>.080</u> 2 nd Alco <u>.080</u>
Signature <u>[Signature]</u>	
(OVER)	

bmb
3/24/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:



- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program, performed by a person trained and qualified to conduct verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 5th day of April, 2010.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130297

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130297

MARCH 11, 2010

OPERATOR'S NAME:

HOELSCHER/JAMES/C

OPERATOR'S NUMBER: 5343

SUBJECT'S LAST NAME:

VERIFICATION

SUBJECT'S FIRST NAME/MI :

OF CALIBRATION

O.L. #: N/A

DEPT/AGENCY: HPB1

CASE/REPORT: N/A

TEST TYPE: U

ALCO TARGET VALUE: .000

ALCO S/N: X301526

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.83 in
ALCO TARGET .079 14:54
BLANK TEST .000 14:55
INTERNAL STANDARD VERIFIED 14:56
ALCO TU 29.83 in .000 14:56
BLANK TEST .000 14:57
SUBJECT SAMPLE .000 14:58
BLANK TEST .000 14:58
ALCO TU 29.83 in .000 14:59
BLANK TEST .000 14:59

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130297

MARCH 11, 2010

TIME 14:53

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 40c
BAROMETER: 29.83 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~||